Allweighs Weighing Services Ltd
Westend Grange
West End Road
Epworth
Doncaster
DN9 1LA
Carried out in partnership with
Shropeshire Scale Company.
ISO: 9001-2015
Certificate No: GB25/000000026

Avery Berkel DEALER

Tel: 01427 875 867
Web: www.allweighs.co.uk

CJS 202600181

Calibration Certificate of Accuracy

Business Name:	C. H. Middleton Ltd
Address:	65-71 Sprotbrough Rd Doncaster DN54 8BW
Contact Name:	Guy Middleton

Location:	Weighbridge	
Manufacturer:	Ian Fellows Lucid	
Serial/Plant No:	20649	
Capacity:	35000	kg
Increment:	20	kg
Frequency:	12	months
Test Date:	26 March 2026	
Next due:	March 2027	

Off-centre Loading Definitive: 6900

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	6900	0			
2	6900	0			
3	6900	0			
4	6900	0			

☑ Four loadcell



☐ Six loadcell



☐ Single loadcell



☐ Eight loadcell



As Found

Measurement Points:	Reading:	Deviation:
0	0	0
4000	4000	0
8000	8000	0
16000	16020	20
23040	23060	20
33040	33080	40
Repeatability		
33040	33080	40

Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
4000	4000	0
8000	8000	0
16000	16000	0
23040	23040	0
33040	33040	0
Repeatability		
33040	33040	0

Notes: Weighbridge found to be within permitted tolerances.

Traceable Weights Used: LAT 117 25/M/2140

Trading Standards Maximum Tolerance: 60 +/- 60 kg

Calibration method and equipment used traceable to national standards.

Engineer Name: CJS

C Scott

AWSL:CC Issue 1 01/01/2025