
POLYWATCH YH-600B INSTALLATION GUIDE

INSIDE THE BOX

						
Main Device	Spo2 Sensor	Respiratory Kit	Effort Belt	Effort Lead Cable	Medical Tape	Nasal Cannula

POLYWATCH STRUCTURE OVERVIEW




INSTALLATION GUIDE

STEP 1



Fasten the ***Effort Belt*** around your chest

STEP 2




Connect ***Effort Lead Cable*** to either Button

STEP 3



Connect the ***other end*** to the black socket on the ***Respiratory Kit***

STEP 4



Attach the ***Respiratory Kit*** to the Effort Belt

STEP 3



Fit in the ***Nasal Cannula*** into your nostrils secure the Nasal Cannula comfortably

STEP 4



Connect the ***other end*** of the Nasal Cannula to the Respiratory Kit
(*turning clockwise to ensure firmly connected*)

AcuraMed
Closer To You

POLYWATCH YH-600B INSTALLATION GUIDE


INSTALLATION GUIDE (Continue)

STEP 5




Connect the ***Respiratory Kit*** into the ***Flow/COM Port***

STEP 6



Wear the ***Main Device*** to your left wrist

STEP 7



Insert the ***Spo2 Sensor*** into your index finger

STEP 8



Switch on the system by pressing down the ***On/OFF/OK Button*** for approximately ***5 sec***

STEP 9



Choose the ***Record function*** by pressing the ***On/Off/OK Button.***
(This will lead to the next screen to check all accessories if connect properly)

STEP 10



Press the ***On/Off/OK button*** for approximately 5 sec to start Recording
(*Ensure run time starts counting on the left top*)



Sleep for minimum 4 to 6 hours for the minimum data collection.

(DO NOT REMOVE THE DEVICE OR ACCESSORY AFTER RECORDING STARTS)

Kindly contact your sales representative if any issue occur during installation

AcuraMed
Closer To You